



MORTAR PRISM FLEXURAL DEVICE

Standard: EN 196/1, EN / ISO 679

Cement Compression & Flexural Testers

For flexural testing on 40.1 x 40 x 160 mm specimens. Made of steel. Two lower bearers with 100 mm span, upper longitudinal self-aligning bearer fitted to ball setting.

Technical Specifications :

Model Number	NL 3027 X / 002 – P 001
Specimen Size	40.1 x 40 x 160 mm
Dimension (mm)	160 (L) x 88 (W) x 169 (H)
Approx. Weight	5.6 kg